


Specification List

Version: 07/June/2022



Allgemeine Hinweise

Dieses Leistungsverzeichnis enthält alle vom Klinisch-immunologischen Labor angebotenen Untersuchungen mit Angaben über die Methoden, die Probenarten und –volumina, die Häufigkeit der Untersuchung, Grenzwerte und ggf. präanalytische Besonderheiten. Das aktuelle Leistungsverzeichnis kann auf unserer Homepage unter www.labor-stoecker.de heruntergeladen werden.

Bearbeitungsdauer:

Die Proben werden am Tag des Eingangs im Laboratorium registriert und die Untersuchungen wenn möglich am gleichen Tag durchgeführt oder begonnen, so dass der Befundbericht in der Regel spätestens am Tag nach dem Probeneingang an den Empfänger übermittelt werden kann. Wenn Untersuchungen nicht täglich durchgeführt werden sowie durch eventuell notwendige Wiederhol- und/oder Verdünnungsanalysen kann sich die Bearbeitungszeit in Einzelfällen verlängern.

Grenzwerte/Referenzbereiche:

Die im Leistungsverzeichnis enthaltenen Grenzwerte entsprechen den Herstellerempfehlungen der verwendeten Testsysteme und dem aktuellen Stand der Wissenschaft und Technik. Durch neue wissenschaftliche Erkenntnisse und technische Fortschritte in der Analysetechnik können sich diese Grenzwerte ändern. Ausschlaggebend für die Befundinterpretation sind die im Befundbericht angeführten Grenzwerte und sonstigen Hinweise.

Messunsicherheit:

Das Klinisch-immunologische Labor führt regelmäßige Kontrollen durch und ergreift kontinuierlich Maßnahmen, um Schwankungen und verfahrensbedingte Abweichungen der Messergebnisse zu minimieren. Aufgrund der vorhandenen Messunsicherheit sind grenzwertige und schwach positive Ergebnisse bzw. Ergebnisse, die in der Nähe der Grenzwerte oder sonstiger medizinischer Entscheidungsgrenzen liegen, vorsichtig zu interpretieren. Auf Anfrage kann für jede quantitative Untersuchung die individuelle Messunsicherheit zur Verfügung gestellt werden. Die meisten der vom Klinisch-immunologischen Labor durchgeführten Untersuchungen (Antikörperdiagnostik) sind jedoch als qualitativ einzustufen.

Vergleichbarkeit von Untersuchungsergebnissen:

Die Ergebnisse verschiedener Testmethoden (z.B. Immunfluoreszenz, ELISA, Blot-Techniken) zum Nachweis der gleichen Antikörper in einer Probe können unterschiedlich ausfallen. Ursache hierfür können unterschiedliche Antigene sein oder die unterschiedliche Immobilisierung des gleichen Antigens auf den verschiedenen Trägermaterialien, die durch Konformationsänderungen des Antigens oder durch sterische Inhibierung die Bindung der Antikörper an das Antigen beeinflussen.

Laboreigene diagnostische Tests (LDT):

Die eingesetzten Untersuchungsverfahren sind üblicherweise CE-gekennzeichnet. Nicht CE-gekennzeichnete Verfahren, die durch das Labor selbst validiert wurden, sind im Leistungsverzeichnis als laboreigene diagnostische Tests (LDT) gekennzeichnet und unterliegen nicht der Akkreditierung des Klinisch-immunologischen Labors.

General Information

The specifications list contains the range of analysis provided by the clinical immunological laboratory and informs you about the methods, sample types and volumes used, the frequency of analysis, reference values and, if necessary, preanalytical aspects. The specifications list is available on the internet at www.labor-stoecker.de.

Duration:

The samples are registered upon receipt in the laboratory. Preferably, analysis is performed or started on the same day to ensure that the report can be sent to the customer generally one day after receipt of the sample at the latest. The time period may be longer in individual cases if the analysis is not carried out daily or retesting of the sample or performing of a dilution series is required.

Reference range:

The reference values given in the specifications list are in accordance with the manufacturer's recommendations for the test system used and with current scientific and technological knowledge. However, reference ranges may change due to new scientific findings and technical developments in analytical techniques. Interpretation of results is always based on the reference ranges and additional remarks which are given on the report.

Measurement deviations:


The clinical immunological laboratory performs regular checks and maintenance to minimise variations and method-related deviations in the measurement results. Due to existing measurement deviations, borderline and weak-positive results or results that lie near to the reference value or other medical decision limits should be interpreted with care. Information about measurement deviations in individual quantitative analyses can be provided upon request. However, most analyses that are performed by the clinical immunological laboratory (antibody diagnostics) are considered qualitative.

Comparability of test results:

Results from different methods (e.g. immunofluorescence, ELISA, blot techniques) for detecting the same antibodies in a sample may differ. This can be due to the use of different antigens or to varying immobilisation of the same antigen to different carrier materials, which can influence the binding of the antibody to the antigen through conformational changes in the antigen or through steric inhibition.

Laboratory developed test (LDT):

The laboratory tests are usually CE marked. Tests that are not CE marked but validated by the laboratory are identified in the specification list as laboratory developed tests (LDT) and are not covered by the accreditation of the clinical immunological laboratory.



Autoantibodies against

21-Hydroxylase IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Acetylcholine-receptor IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA, IFT
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Acetylcholine receptors


sample type: serum
sample volume: 1,5 ml
method: RIA
frequency: biweekly
reference value: see result report
as at: 12/21

Aggrin IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
Remark: The analysis is a LDT and not covered by the accreditation of our laboratory.
as at: 06/22

Actin (F-Actin) IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21



Autoantibodies against

Actin IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Alveolar basement membrane IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
remark: The analysis is not covered by the accreditation of our laboratory.
as at: 05/22

AMA IgG


sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

AMA M2-3E IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

AMA Profil IgGM (M2-3e, M2, M4, M9)

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: EUROLINE
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21



Autoantibodies against

AMA-M4 IgGM

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: EUROASSAY
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

AMA-M7 (myocard specific) IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

AMA-M9 IgGM


sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: EUROASSAY
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Amphiphysin IgG

sample type: serum, EDTA, heparin, citrat plasma
sample type 2: CSF (LDT, not accredited)
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 05/22

ANA IgG (autoantibodies against nuclei) IF global test

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21



Autoantibodies against

ANA Profile 23 IgG (nukleosomes, dsDNA, histones, SS-A, Ro-52, SS-B, nRNP/Sm, Sm, Mi alpha, Mi beta, Ku, CENP A, CENP B, Sp100, PML, Scl-70, Pm-Scl100, Pm-Scl75, RP11, RP155, gp210, PCNA, DFS70)

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: EUROLINE
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

ANCA (autoantibodies against granulocyte cytoplasm)

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

ANCA IgG (MPO, PR3)


sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: EUROLINE
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

ANCA Profile IgG (PR3, MPO, Elastase, Cath.G, BPI, Lactoferrin)

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

ANNA-3 IgG

sample type: serum, EDTA, heparin, citrat plasma
sample type 2: CSF (LDT, not accredited)
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 05/22



Autoantibodies against

Anti Neurofilaments IgG

sample type: serum, EDTA, heparin, citrat plasma
sample type 2: CSF (LDT, not accredited)
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 06/22

Aquaporin-4 IgG

sample type: serum, EDTA, heparin, citrat plasma
sample type2: CSF
sample volume: 2,0 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

ASCA (Sacch. cerevisiae) IgA


sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

ASCA (Sacch. cerevisiae) IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Asialoglycoprotein rec. (ASGPR) IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
Remark: The analysis is a LDT and not covered by the accreditation of our laboratory.
as at: 05/22



Autoantibodies against

ASMA (GMA) IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

AT1A3 IgG

sample type: serum, EDTA, heparin, citrat plasma
sample type 2: CSF
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
remark: The analysis is a LDT and not covered by the accreditation of our laboratory.
as at: 06/22

Eye muscle IgAGM


sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 05/22

Autoantibody Profile IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
remark: The analysis is a LDT and not covered by the accreditation of our laboratory.
as at: 05/22

Basal ganglia IgG

sample type: serum, EDTA, heparin, citrat plasma
sample type 2: CSF
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
remark: The analysis is a LDT and not covered by the accreditation of our laboratory.
as at: 05/22



Autoantibodies against

GABA-a-rezeptors IgG

sample type: serum, EDTA, heparin, citrat plasma

sample type 2: CSF

sample volume: 1,5 ml

method: IFA

frequency: daily (mo-fri)

reference value: see result report

remark: The analysis is a LDT and not covered by the accreditation of our laboratory.

as at: 06/22

Goblet cells intestinal IgA

sample type: serum

sample type 2: EDTA, heparin, citrat plasma

sample volume: 1,5 ml

method: IFA

frequency: daily (mo-fri)

reference value: see result report

as at: 12/21

Goblet cells intestinal IgG

sample type: serum

sample type 2: EDTA, heparin, citrat plasma

sample volume: 1,5 ml

method: IFA

frequency: daily (mo-fri)

reference value: see result report

as at: 12/21

BP180 (NC16A-4x) IgG

sample type: serum

sample type 2: EDTA, heparin, citrat plasma

sample volume: 1,5 ml

method: ELISA, IFA

frequency: daily (mo-fri)

reference value: see result report

as at: 12/21

BP230 IgG

sample type: serum

sample type 2: EDTA, heparin, citrat plasma


sample volume: 1,5 ml

method: ELISA, IFA

frequency: daily (mo-fri)

reference value: see result report

as at: 12/21



Autoantibodies against

C1q IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Calcium channels PQ-type IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: RIA
frequency: biweekly
reference value: see result report
as at: 12/21

cANCA IgG


sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Cardiolipin (AMA-M1) IgAGM

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Cardiolipin IgA

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21



Autoantibodies against

Cardiolipin IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Cardiolipin IgM

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

CARPVIII IgG


sample type: serum, EDTA, heparin, citrat plasma
sample type 2: CSF
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
remark: The analysis is a LDT and not covered by the accreditation of our laboratory.
as at: 06/22

CASPR1 IgG

sample type: serum, EDTA, heparin, citrat plasma
sample type 2: CSF
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
remark: The analysis is a LDT and not covered by the accreditation of our laboratory.
as at: 06/22

CASPR2 IgG

sample type: serum, EDTA, heparin, citrat plasma
sample type 2: CSF
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21



Autoantibodies against

Cathepsin G IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

CCP (cycl. citrull. peptide) IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

CEP-1 IgG


sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

cN-1A (Mup44) IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Coilin; P80 (Few Nuclear Dots) IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21



Autoantibodies against

Contactin1 IgG

sample type: serum, EDTA, heparin, citrat plasma

sample type 2: CSF

sample volume: 1,5 ml

method: IFA

frequency: daily (mo-fri)

reference value: see result report

remark: The analysis is a LDT and not covered by the accreditation of our laboratory.

as at: 06/22

Cornea IgAGM

sample type: serum

sample type 2: EDTA, heparin, citrat plasma

sample volume: 1,5 ml

method: IFA

frequency: daily (mo-fri)

reference value: see result report

as at: 05/22

Corpus ciliare IgAGM

sample type: serum

sample type 2: EDTA, heparin, citrat plasma

sample volume: 1,5 ml

method: IFA

frequency: daily (mo-fri)

reference value: see result report

as at: 05/22

CV2 (CRMP-5) IgG

sample type: serum, EDTA, heparin, citrat plasma

sample type 2: CSF (LDT, not accredited)

sample volume: 1,5 ml

method: IFA

frequency: daily (mo-fri)

reference value: see result report

as at: 05/22

CV2 IgG

sample type: serum, EDTA, heparin, citrat plasma

sample type 2: CSF


sample volume: 1,5 ml

method: EUROLINE

frequency: daily (mo-fri)

reference value: see result report

as at: 12/21



Autoantibodies against

CXCL13

sample type: CSF
sample volume: 1,5 ml
method: ELISA
frequency: weekly
reference value: see result report
as at: 01/19

Cyclin I (PCNA) IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Cytoplasm HEp-2-cells IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Cytoplasm Profile 13 Ag IgG (M2, M2-3e, rib.P-Prot., Jo-1, SRP, PL-7, PL-12, EJ, OJ, Ha, Ks, Zo, Ro52)


sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: EUROLINE
frequency: daily (mo-fri)
reference value: see result report
as at: 05/22

Desmoglein 1 IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA, IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Desmoglein 3 IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA, IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21



Autoantibodies against

Desmosomes (intercellular substance) IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

DFS70 IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

DNA bound lactoferrin IgA


sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
remark: The analysis is a LDT and not covered by the accreditation of our laboratory.
as at: 05/22

DNA bound lactoferrin IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Dopamin-receptors 2 IgG

sample type: serum, EDTA, heparin, citrat plasma
sample type 2: CSF
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
remark: The analysis is a LDT and not covered by the accreditation of our laboratory.
as at: 06/22



Autoantibodies against

DPPX IgG

sample type: serum, EDTA, heparin, citrat plasma
sample type 2: CSF (LDT, not accredited)
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 05/22

dsDNA IgA (Crithidia luc.)

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
remark: The analysis is a LDT and not covered by the accreditation of our laboratory.
as at: 05/22

dsDNA IgAGM


sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: RIA
frequency: biweekly
reference value: see result report
as at: 12/21

dsDNA IgG (Crithidia luc.)

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

dsDNA IgM (Crithidia luc.)

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
remark: The analysis is a LDT and not covered by the accreditation of our laboratory.
as at: 05/22



Autoantibodies against

dsDNS-NcX IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

EJ IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: EUROLINE
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Elastase IgG


sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Elastin IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

ENA Pool Plus IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21



Autoantibodies against

ENA Profile Plus 1 IgG (nRNP/Sm, Sm, SS-A, SS-B, Scl-70, Jo-1)

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

ENA Profile Plus IgG EUROASSAY (nRNP/Sm, Sm, SS-A, SS-B, Scl-70, Jo-1)

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: EUROASSAY
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

ENA SLE Profile 2 (dsDNA, histones, nukleosomes, nRNP/Sm, Sm, SS-A, SS-B, Scl-70)


sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Endomysium IgA

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Endomysium IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21



Autoantibodies against

Enterocytes IgA

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
remark: The analysis is a LDT and not covered by the accreditation of our laboratory.
as at: 05/22

Enterocytes IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
remark: The analysis is a LDT and not covered by the accreditation of our laboratory.
as at: 05/22

Envoplakin IgG


sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Epidermal basement membrane IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

ERC1 IgG

sample type: serum, EDTA, heparin, citrat plasma
sample type 2: CSF
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
remark: The analysis is a LDT and not covered by the accreditation of our laboratory.
as at: 06/22



Autoantibodies against

Filaggrin (RA keratine) IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
remark: The analysis is a LDT and not covered by the accreditation of our laboratory.
as at: 05/22

Flotillin 1/2 IgG

sample type: serum, EDTA, heparin, citrat plasma
sample type 2: CSF
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
remark: The analysis is a LDT and not covered by the accreditation of our laboratory.
as at: 06/22

GABA-a-rezeptors IgG


sample type: serum, EDTA, heparin, citrat plasma
sample type 2: CSF
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
remark: The analysis is a LDT and not covered by the accreditation of our laboratory.
as at: 06/22

Rho-GTPase activating protein 26 IgG

sample type: serum, EDTA, heparin, citrat plasma
sample type 2: CSF
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
remark: The analysis is a LDT and not covered by the accreditation of our laboratory.
as at: 06/22

GABA-b-receptors IgG

sample type: serum, EDTA, heparin, citrat plasma
sample type 2: CSF
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21



Autoantibodies against

GAD 65 kDa IgG

sample type: serum, EDTA, heparin, citrat plasma
sample type 2: CSF (LDT, not accredited)
sample volume: 1,5 ml
method: ELISA, IFA
frequency: ELISA biweekly, IFA daily (mo-fri)
reference value: see result report
as at: 05/22

GAD IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Bile canaliculi IgG


sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
remark: The analysis is a LDT and not covered by the accreditation of our laboratory.
as at: 05/22

Bile duct IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
remark: The analysis is a LDT and not covered by the accreditation of our laboratory.
as at: 05/22

Bile duct epithelium IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
remark: The analysis is a LDT and not covered by the accreditation of our laboratory.
as at: 05/22



Autoantibodies against

Ganglioside Profile 2 IgG (GM1, GM2, GM3, GD1a, GD1b, GT1b, GQ1b)

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: EUROLINE
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Ganglioside Profile 2 IgM (GM1, GM2, GM3, GD1a, GD1b, GT1b, GQ1b)

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: EUROLINE
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

GBM (glomerular basement membrane) IgG


sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA, IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Vascular endothelium IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 05/22

Total Tau

sample type: CSF
sample volume: 2,0 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
remark: Please note the instructions for preanalytics. The samples must be submitted frozen.
remark 2: This analysis is not covered by the current accreditation scope.
as at: 01/19



Autoantibodies against

GFAP (Glial fibrillary acidic protein) IgG

sample type: serum, EDTA, heparin, citrat plasma

sample type 2: CSF

sample volume: 1,5 ml

method: IFA

frequency: daily (mo-fri)

reference value: see result report

remark: The analysis is a LDT and not covered by the accreditation of our laboratory.

as at: 05/22

Glia nuclear antibodies (AGNA) IgG

sample type: serum, EDTA, heparin, citrat plasma

sample type 2: CSF (LDT, not accredited)

sample volume: 1,5 ml

method: IFA

frequency: daily (mo-fri)

reference value: see result report

as at: 05/22

Gliadin-GAF-3X (Z-AGFA) IgA

sample type: serum

sample type 2: EDTA, heparin, citrat plasma

sample volume: 1,5 ml

method: ELISA

frequency: daily (mo-fri)

reference value: see result report

as at: 12/21

Gliadin-GAF-3X (Z-AGFA) IgG

sample type: serum

sample type 2: EDTA, heparin, citrat plasma

sample volume: 1,5 ml

method: ELISA

frequency: daily (mo-fri)

reference value: see result report

as at: 12/21

GluRD2 IgG

sample type: serum, EDTA, heparin, citrat plasma

sample type 2: CSF

sample volume: 1,5 ml


method: IFA

frequency: daily (mo-fri)

reference value: see result report

remark: The analysis is a LDT and not covered by the accreditation of our laboratory.

as at: 06/22



Autoantibodies against

Glutamate receptor (type AMPA) IgG

sample type: serum, EDTA, heparin, citrat plasma
sample type 2: CSF (LDT, not accredited)
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 05/22

Glutamate receptor (Typ NMDA) IgG

sample type: serum, EDTA, heparin, citrat plasma
sample type 2: CSF
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Glycine receptors IgG


sample type: serum, EDTA, heparin, citrat plasma
sample type 2: CSF
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
remark: The analysis is a LDT and not covered by the accreditation of our laboratory.
as at: 05/22

Golgi apparatus IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Heart muscle IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21



Autoantibodies against

Histones global IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Homer 3 IgG

sample type: serum, EDTA, heparin, citrat plasma
sample type 2: CSF
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
remark: The analysis is a LDT and not covered by the accreditation of our laboratory.
as at: 06/22

Hu (ANNA-1) IgG


sample type: serum, EDTA, heparin, citrat plasma
sample type 2: CSF (LDT, not accredited)
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 05/22

Pituitary gland posterior lobe IgAGM

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Pituitary gland anterior lobe IgAGM

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21



Autoantibodies against

IA2 (Tyrosinphosphatase) IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: biweekly
reference value: see result report
as at: 12/21

ICA (islet cells) IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-thu)
reference value: see result report
as at: 12/21

IgLON5 IgG

sample type: serum, EDTA, heparin, citrat plasma
sample type 2: CSF (LDT, not accredited)
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Insulin human IgG


sample type: serum
sample volume: 1,5 ml
method: RIA
frequency: weekly
reference value: see result report
as at: 12/21

Intrinsic factor IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

ITPR1 IgG

sample type: serum, EDTA, heparin, citrat plasma
sample type 2: CSF
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
remark: The analysis is a LDT and not covered by the accreditation of our laboratory.
as at: 06/22



Autoantibodies against

Jo-1 IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Potassium channels

sample type: serum
sample type 2: CSF (LDT, not accredited)
sample volume: 1,5 ml
method: RIA
frequency: weekly
reference value: see result report
as at: 05/22

KCNA2 IgG


sample type: serum, EDTA, heparin, citrat plasma
sample type 2: CSF
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
remark: The analysis is a LDT and not covered by the accreditation of our laboratory.
as at: 06/22

Nuclear membrane IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Collagen type VII IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 04/22



Autoantibodies against

Laminin 5 (332) IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Laminin IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

LC-1 (Liver cytosol) IgG


sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Liver IgG IFA (globaltest, 6 biochips)

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Liver Profile (ALD) IgG (AMA-M2, M2-3e (BPO), Sp100, PML, gp210, LC-1, LKM-1, SLA/LP, SS-A, Ro-52, Scl-70, CENP A, CENP B, PGDH)

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: EUROLINE
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21



Autoantibodies against

Liver Profile (AMA-M2, LKM-1, LC-1, SLA/LP)

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: EUROASSAY
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Lens oculi IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

LGI1 IgG


sample type: serum, EDTA, heparin, citrat plasma
sample type 2: CSF (LDT, not accredited)
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 05/22

LKM-1 IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA, ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

LKM-2 IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
remark: The analysis is a LDT and not covered by the accreditation of our laboratory.
as at: 05/22



Autoantibodies against

LMA (Liver cell membrane) IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
remark: The analysis is a LDT and not covered by the accreditation of our laboratory.
as at: 05/22

LRP 4 IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
Remark: The analysis is a LDT and not covered by the accreditation of our laboratory.
as at: 06/22

LSP (Liver-spec. proteine) IgG


sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Lysosomes IgG

sample type: serum
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/16

Ma (Ma1, Ma2/Ta) IgG

sample type: serum, EDTA, heparin, citrat plasma
sample type 2: CSF (LDT)
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 05/22



Autoantibodies against

MAG (Myelin assoz. Glykopr.) IgA

sample type: serum, EDTA, heparin, citrat plasma

sample type 2: CSF

sample volume: 1,5 ml

method: IFA

frequency: daily (mo-fri)

reference value: see result report

remark: The analysis is a LDT and not covered by the accreditation of our laboratory.

as at: 05/22

MAG (Myelin assoz. Glykopr.) IgG

sample type: serum, EDTA, heparin, citrat plasma

sample type 2: CSF (LDT, not accredited)

sample volume: 1,5 ml

method: IFA

frequency: daily (mo-fri)

reference value: see result report

as at: 05/22

MAG (Myelin assoz. Glykopr.) IgM

sample type: serum, EDTA, heparin, citrat plasma

sample type 2: CSF (LDT, not accredited)

sample volume: 1,5 ml

method: IFA

frequency: daily (mo-fri)

reference value: see result report

as at: 05/22

MBP (Myelobas. protein) IgG

sample type: serum, EDTA, heparin, citrat plasma

sample type 2: CSF

sample volume: 1,5 ml

method: IFA

frequency: daily (mo-fri)

reference value: see result report

remark: The analysis is a LDT and not covered by the accreditation of our laboratory.

as at: 05/22

MBP (Myelobas. protein) IgM

sample type: serum, EDTA, heparin, citrat plasma

sample type 2: CSF

sample volume: 1,5 ml

method: IFA

frequency: daily (mo-fri)

reference value: see result report

remark: The analysis is a LDT and not covered by the accreditation of our laboratory.

as at: 06/22



Autoantibodies against

Melanocytes IgAGM

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

mGluR1 (metabotropic glutamate receptor 1) IgG

sample type: serum, EDTA, heparin, citrat plasma
sample type 2: CSF
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
remark: The analysis is a LDT and not covered by the accreditation of our laboratory.
as at: 05/22

mGluR5 (metabotropic glutamate receptor 5) IgG


sample type: serum, EDTA, heparin, citrat plasma
sample type 2: CSF
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
remark: The analysis is a LDT and not covered by the accreditation of our laboratory.
as at: 05/22

Mi-2 IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Mitosin (CENP-F) IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21



Autoantibodies against

MOG (Myelin-oligodendrocytes glycoprotein) IgG

sample type: serum, EDTA, heparin, citrat plasma
sample type 2: CSF (LDT, not accredited)
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 05/22

MPO (Myeloperoxidase) IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

MPO/PR3/GBM Profile IgG


sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: EUROLINE
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

MSA-1 (NuMa-1) IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

MSA-2 (HsEg5, spindle apparatus) IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21



Autoantibodies against

MSA-3 (CENP-F chromosomes ass. AG) IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

MuSK (musclespecific tyrosinkinase) IgG

sample type: serum
sample volume: 1,5 ml
method: RIA
frequency: biweekly
reference value: see result report
as at: 12/21

MuSK IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFT
frequency: daily (Mo-Fr)
reference value: see result report
as at: 12/21

muscle intercalated disk IgG


sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Myelin IgG

sample type: serum, EDTA, heparin, citrat plasma
sample type 2: CSF (LDT, not accredited)
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 05/22

Myolemma IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
remark: The analysis is a LDT and not covered by the accreditation of our laboratory.
as at: 05/22



Autoantibodies against

Myopathie Profile 4 IgG (Mi-2alpha/beta, TIF1gamma, MDA5, NXP2, SAE1, Ku, PM100, PM75, Jo-1, SRP, PL-7, PL12, EJ, OJ, Ro52, cN-1a, Ha, Ks, Zo)

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: EUROLINE
frequency: daily (mo-fri)
reference value: see result report
as at: 05/22

Myosin IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Adrenal cortex IgG


sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Parathyroid gland IgAGM

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Neuronal antigens IgG

sample type: serum, EDTA, heparin, citrat plasma
sample type 2: CSF (LDT, not accredited)
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 05/22



Autoantibodies against

Neurexin-3-alpha IgG

sample type: serum, EDTA, heparin, citrat plasma

sample type 2: CSF

sample volume: 1,5 ml

method: IFA

frequency: daily (mo-fri)

reference value: see result report

remark: The analysis is a LDT and not covered by the accreditation of our laboratory.

as at: 06/22

Neurochondrin IgG

sample type: serum, EDTA, heparin, citrat plasma

sample type 2: CSF

sample volume: 1,5 ml

method: IFA

frequency: daily (mo-fri)

reference value: see result report

remark: The analysis is a LDT and not covered by the accreditation of our laboratory.

as at: 06/22

Neurofascin 155 IgG

sample type: serum, EDTA, heparin, citrat plasma

sample type 2: CSF

sample volume: 1,5 ml

method: IFA

frequency: daily (mo-fri)

reference value: see result report

remark: The analysis is a LDT and not covered by the accreditation of our laboratory.

as at: 06/22

Neurofascin 186 IgG

sample type: serum, EDTA, heparin, citrat plasma

sample type 2: CSF

sample volume: 1,5 ml

method: IFA

frequency: daily (mo-fri)

reference value: see result report

remark: The analysis is a LDT and not covered by the accreditation of our laboratory.

as at: 06/22

Neurofilament (pNf-H) ELISA

sample type: serum

sample type 2: CSF

sample volume: 1,5 ml

method: ELISA


frequency: daily (mo-fri)

reference value: see result report

remark: Please note the instructions for preanalytics. The samples must be submitted frozen.

remark 2: This analysis is not covered by the current accreditation scope.

as at: 01/19



Autoantibodies against

Neuronale immunoblot IgG (amphiphysin, CV2, PNMA2 (Ma2/Ta), Ri, Yo, Hu, Recoverin, SOX1, Titin, Zic4, GAD65, Tr (DNER))

sample type: serum, EDTA, heparin, citrat plasma
sample type 2: CSF
sample volume: 1,5 ml
method: EUROLINE
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

NF-light (Neurofilament light) ELISA

sample type: CSF
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
remark: Please note the instructions for preanalytics. The samples must be submitted frozen.
remark 2: This analysis is not covered by the current accreditation scope.
as at: 05/2020

Kidney IgG


sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

NOR (nucleus-organisator-region) 90 IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Nucleosomes IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21



Autoantibodies against

OJ IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: EUROLINE
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Ovary IgAGM

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Ovary tecacells IgAGM


sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Ovary zona pellucida IgAGM

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

pANCA IgA

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21



Autoantibodies against

pANCA IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Pancreas execution duct IgA

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
remark: The analysis is a LDT and not covered by the accreditation of our laboratory.
as at: 05/22

Pancreas execution duct IgG


sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
remark: The analysis is a LDT and not covered by the accreditation of our laboratory.
as at: 05/22

Pancreas acini IgA (CUZD1, GP2)

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Pankreas acini IgG (CUZD1, GP2)

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21



Autoantibodies against

PCA (H/K-ATPase) IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

PCA (parietalcells) IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

PCA-2 (purkinjecell-cytopl.) IgG


sample type: serum, EDTA, heparin, citrat plasma
sample type 2: CSF (LDT)
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 05/22

PDGH IgG EUROLINE

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: EUROLINE
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Phosphatidyl-serin IgA

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21



Autoantibodies against

Phosphatidyl-serin IgAGM

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Phosphatidyl-serin IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Phosphatidyl-serin IgM


sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Phospholipase A2-receptor IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA, ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

PL-12 IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: EUROLINE
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21



Autoantibodies against

PL-7 IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: EUROLINE
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Plexus myentericus IgG

sample type: serum, EDTA, heparin, citrat plasma
sample type 2: CSF (LDT, not accredited)
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 05/22

PM-1/PM-Sci IgG


sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

PR3-hn-hr (proteinase3) IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

pTau(181)

sample type: CSF
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
remark: Please note the instructions for preanalytics. The samples must be submitted frozen.
remark 2: This analysis is not covered by the current accreditation scope.
as at: 01/19



Autoantibodies against

RANA (Rheumatoide arthritis nuclear antigen) IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Retina IgAGM

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Rheumatoid factor (class. Rheumatoid factor) IgM


sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Rheumatoid factor IgA

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Rheumatoid factor IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21



Autoantibodies against

Rho-GTPase activating protein 26 IgG

sample type: serum, EDTA, heparin, citrat plasma

sample type 2: CSF

sample volume: 1,5 ml

method: IFA

frequency: daily (mo-fri)

reference value: see result report

remark: The analysis is a LDT and not covered by the accreditation of our laboratory.

as at: 06/22

Ri (ANNA-2) IgG

sample type: serum, EDTA, heparin, citrat plasma

sample type 2: CSF (LDT, not accredited)

sample volume: 1,5 ml

method: IFA

frequency: daily (mo-fri)

reference value: see result report

as at: 05/22

Ribosomal P-proteins IgG

sample type: serum

sample type 2: EDTA, heparin, citrat plasma

sample volume: 1,5 ml

method: ELISA

frequency: daily (mo-fri)

reference value: see result report

as at: 05/22

Ribosomes IgG

sample type: serum

sample type 2: EDTA, heparin, citrat plasma

sample volume: 1,5 ml

method: IFA

frequency: daily (mo-fri)

reference value: see result report

as at: 12/21

RNA-polymerase IgG

sample type: serum

sample type 2: EDTA, heparin, citrat plasma


sample volume: 1,5 ml

method: IFA

frequency: daily (mo-fri)

reference value: see result report

as at: 12/21



Autoantibodies against

Ro 52 IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: EUROLINE
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Sa IgG

sample type: serum
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Sarcolemma IgG


sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
remark: The analysis is a LDT and not covered by the accreditation of our laboratory.
as at: 05/22

Thyroid microsomes IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Sci-70 (DNS topoisomerase I) IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21



Autoantibodies against

Signal recognition particle (SRP) IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: EUROLINE
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Muscle striated IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

SLA/LP IgG


sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Sm IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

SOX-1 IgG

sample type: serum, EDTA, heparin, citrat plasma
sample type 2: CSF
sample volume: 1,5 ml
method: EUROLINE
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21



Autoantibodies against

Sp100, PML (nuclear dots) IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

saliva gland execution ducts IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

saliva gland acini IgG


sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Salivary gland IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Spermatozoa IgA

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21



Autoantibodies against

Spermatozoa IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Spermatozoa IgM

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

SS-A (Ro) 60 kDA


sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

SS-B (La) IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

ssDNS IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21



Autoantibodies against

Steroidhormonprod. cells (ovary) IgAGM

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 05/22

Systemsclerosis IgG EUROLINE (Sci-70, CENP A, CENP B, RP11, RP155, Fibrillarin, NOR90, Th/To, PM-Sci100, PM-Sci75, Ku, PDGFR, Ro-52)

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: EUROLINE
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Testis IgAGM


sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

TG (thyroglobulin) IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 05/22

Thrombocytes (free Ab) IFA IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21



Autoantibodies against

THSD7A IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
remark:
as at: 12/21

Thymus IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
remark: The analysis is a LDT and not covered by the accreditation of our laboratory.
as at: 05/22

Thyreoglobulin IgG


sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Th/To IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: EUROLINE
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Titin IgG

sample type: serum, EDTA, heparin, citrat plasma
sample type 2: CSF
sample volume: 1,5 ml
method: EUROLINE
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21



Autoantibodies against

To (7-2 RNP) IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: EUROLINE
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

TPO (thyreoperoxidase) IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Tr (purkinjecell cytoplasm) IgG

sample type: serum, EDTA, heparin, citrat plasma
sample type 2: CSF (LDT, not accredited)
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 05/22

TRAk (TSH-receptor) IgG


sample type: serum
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Transglutaminase IgA

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Transglutaminase IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21



Autoantibodies against

Tropomyosin IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Lacrimal gland IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
remark: The analysis is a LDT and not covered by the accreditation of our laboratory.
as at: 05/22

Tubular basement membrane IgG


sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Tunica choroidea IgAGM

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

U1-(n) RNP/Sm IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21



Autoantibodies against

U3-nRNP (Fibrillarin) IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA, EUROLINE
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Vasopressin prod. cells IgAGM

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
remark: The analysis is a LDT and not covered by the accreditation of our laboratory.
as at: 05/22

Vimentin IgG


sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Vinculin IgAGM

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
remark: The analysis is a LDT and not covered by the accreditation of our laboratory.
as at: 05/22

Yo (PCA-1) IgG

sample type: serum, EDTA, heparin, citrat plasma
sample type 2: CSF
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21



Autoantibodies against

Centrioles IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Centromere protein B IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Centromeres IgG


sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Zic4 IgG

sample type: serum, EDTA, heparin, citrat plasma
sample type 2: CSF (LDT, not accredited)
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 05/22

ZIC4 IgG

sample type: serum, EDTA, heparin, citrat plasma
sample type 2: CSF
sample volume: 1,5 ml
method: EUROLINE
frequency: daily (mo-fri)
reference value: see result report
as at: 05/22



Autoantibodies against

Zinc transporter (ZnT8)

sample type: serum
sample type 2: heparin, citrate plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Circulating immun complexes C1q

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: weekly
reference value: see result report
as at: 12/21

Cytoskeleton IgG


sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Coeliac disease Profile IgA

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: EUROLINE
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Coeliac disease Profile IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: EUROLINE
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21



Autoantibodies against

β-amyloid (1-40) ELISA

sample type: CSF
sample volume: 2,0 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
remark: Please note the instructions for preanalytics. The samples must be submitted frozen.
remark 2: This analysis is not covered by the current accreditation scope.
as at: 01/19

β-amyloid (1-42) ELISA

sample type: CSF
sample volume: 2,0 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
remark: Please note the instructions for preanalytics. The samples must be submitted frozen.
remark 2: This analysis is not covered by the current accreditation scope.
as at: 01/19

β2-glycoprotein IgA


sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

β2-glycoprotein IgAGM

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

β2-glycoprotein IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21



Autoantibodies against

β2-glycoprotein IgM

sample type: serum

sample type 2: EDTA, heparin, citrat plasma


sample volume: 1,5 ml

method: ELISA

frequency: daily (mo-fri)

reference value: see result report

as at: 12/21



Antibodies against

Adenovirus IgA

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 03/22

Adenovirus IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 03/22

Adenovirus IgM


sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 03/22

Bartonella henselae IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Bartonella henselae IgM

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21



Antibodies against

Bartonella quintana IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Bartonella quintana IgM

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Bordetella parapertussis IgA


sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
remark: The analysis is a LDT and not covered by the accreditation of our laboratory.
as at: 05/22

Bordetella parapertussis IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Bordetella parapertussis IgM

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
remark: The analysis is a LDT and not covered by the accreditation of our laboratory.
as at: 05/22



Antibodies against

Bordetella pertussis IgA

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
remark: Bei dieser Untersuchung handelt sich es um einen LDT, sie unterliegt nicht der Akkreditierung.
as at: 05/22

Bordetella pertussis IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Bordetella pertussis IgM


sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA, ELISA
frequency: daily (mo-fri)
reference value: see result report
remark: The analysis is a LDT and not covered by the accreditation of our laboratory.
as at: 05/22

Bordetella-FHA IgA

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Bordetella-FHA IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21



Antibodies against

Bordetella pertussis IgA

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: EUROLINE
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Bordetella-pertussis IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: EUROLINE
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Bordetella-pertussis IgM


sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Bordetella-pertussis-Toxin IgA

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Bordetella-pertussis-Toxin IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21



Antibodies against

Borrelia afzelii IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA, Westernblot
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Borrelia afzelii IgM

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: Westernblot
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Borrelia burgdorferi (USA) IgG


sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA, Westernblot
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Borrelia burgdorferi (USA) IgM

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA, Westernblot
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Borrelia garinii IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA, Westernblot
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21



Antibodies against

Borrelia garinii IgM

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA, Westernblot
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Borrelia IgM

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA, EUROLINE/Westernblot
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Borrelia CSF/serum quotient IgG


sample type: serum
sample type 2: CSF
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Borrelia CSF/serum quotient IgM

sample type: serum
sample type 2: CSF
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Borrelia plus VisE IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21



Antibodies against

Borrelia-RN-AT IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: EUROLINE
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Borrelia-RN-AT IgM

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: EUROLINE
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Brucella abortus IgG


sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Brucella abortus IgM

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Campylobacter coli IgA

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
remark: The analysis is a LDT and not covered by the accreditation of our laboratory.
as at: 05/22



Antibodies against

Campylobacter coli IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
remark: The analysis is a LDT and not covered by the accreditation of our laboratory.
as at: 05/22

Campylobacter jejuni IgA

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA, ELISA
frequency: daily (mo-fri)
reference value: see result report
remark: The analysis is a LDT and not covered by the accreditation of our laboratory.
as at: 05/22

Campylobacter jejuni IgG


sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA, ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Candida albicans IgA

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Candida albicans IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21



Antibodies against

Candida albicans IgM

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Candida crusei IgAGM

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
remark: The analysis is a LDT and not covered by the accreditation of our laboratory.
as at: 05/22

Candida glabrata IgAGM


sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
remark: The analysis is a LDT and not covered by the accreditation of our laboratory.
as at: 05/22

Candida parapsilosis IgAGM

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
remark: The analysis is a LDT and not covered by the accreditation of our laboratory.
as at: 05/22

Candida tropicalis IgAGM

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
remark: The analysis is a LDT and not covered by the accreditation of our laboratory.
as at: 05/22



Antibodies against

Chikungunya Virus IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: biweekly
reference value: see result report
as at: 12/21

Chikungunya Virus IgM

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: biweekly
reference value: see result report
as at: 12/21

Chikungunya virus IgG


sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: biweekly
reference value: see result report
as at: 12/21

Chikungunya Virus IgM

sample type: serum
sample type 2: EDTA, heparin plasma
sample volume: 1,5 ml
method: ELISA
frequency: biweekly
reference value: see result report
as at: 12/21

Chlamydia pneumoniae IgA

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA, IFA, MIF
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21



Antibodies against

Chlamydia pneumoniae IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA, IFA, MIF
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Chlamydia pneumoniae IgM

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA, IFA, MIF
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Chlamydia psittaci IgA MIF


sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: MIF
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Chlamydia psittaci IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: MIF
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Chlamydia psittaci IgM

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: MIF
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21



Antibodies against

Chlamydia trachomatis IgA

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: EUROLINE/Westernblot
frequency: daily (mo-fri)
reference value: see result report
remark: The analysis is not covered by the accreditation of our laboratory.
as at: 05/22

Chlamydia trachomatis IgA

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA, IFA, MIF
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Chlamydia trachomatis IgG


sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: EUROLINE/Westernblot
frequency: daily (mo-fri)
reference value: see result report
remark: The analysis is not covered by the accreditation of our laboratory.
as at: 05/22

Chlamydia trachomatis IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA, IFA, MIF
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Chlamydia trachomatis IgM

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA, IFA, MIF
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21



Antibodies against

CMV IgG avidity

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

CMV IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
remark: This analysis is not intended to assess suitability for transfusion, transplantation or cell therapy.
as at: 05/22

CMV IgM


sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
remark: This analysis is not intended to assess suitability for transfusion, transplantation or cell therapy.
as at: 05/22

CMV CSF/serum quotient IgG

sample type: serum
sample type 2: CSF
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Coxsackie A IgA

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21



Antibodies against

Coxsackie A IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Coxsackie A IgM

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Coxsackie B IgA


sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Coxsackie B IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Coxsackie B IgM

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21



Antibodies against

CXCL13

sample type: CSF
sample volume: 1,5 ml
method: ELISA
frequency: weekly
reference value: see result report
as at: 01/19

Dengue Virus 1-4 IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA, IFA
frequency: biweekly
reference value: see result report
as at: 12/21

Dengue Virus 1-4 IgM

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA, IFA
frequency: biweekly
reference value: see result report
as at: 12/21

Dengue Virus NS1 ELISA


sample type: serum
sample type 2: EDTA, heparin plasma
sample volume: 1,5 ml
method: ELISA
frequency: biweekly
reference value: see result report
as at: 12/21

Diphtherie Toxoid IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

EBNA IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA, IFA
frequency: daily (mo-fri)
reference value: see result report
remark: This analysis is not intended to assess suitability for transfusion, transplantation or cell therapy.
as at: 05/22



Antibodies against

EBV-CA IgA

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA, IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

EBV-CA IgG Avidity

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA, IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

EBV-CA IgG


sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA, IFA
frequency: daily (mo-fri)
reference value: see result report
remark: This analysis is not intended to assess suitability for transfusion, transplantation or cell therapy.
as at: 05/22

EBV-CA IgM

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA, IFA
frequency: daily (mo-fri)
reference value: see result report
remark: This analysis is not intended to assess suitability for transfusion, transplantation or cell therapy.
as at: 05/22

EBV-EA IgA

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA, IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21



Antibodies against

EBV-EA IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA, ELISA
frequency: daily (mo-fri)
reference value: see result report
remark: This analysis is not intended to assess suitability for transfusion, transplantation or cell therapy.
as at: 05/22

EBV-EA IgM

sample type: serum
sample type 2: EDTA plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

EBV-Profile 2 EUROLINE IgG (VCA gp125, VCA p19, EBNA-1, p22, EA-D)


sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: EUROLINE
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

EBV-Profile 2 EUROLINE IgM (VCA gp125, VCA p19, EBNA-1, p22, EA-D)

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: EUROLINE
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Echinococcus granulosus IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA, ELISA, EUROLINE/Westernblot
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21



Antibodies against

Echinococcus granulosus IgM

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

TBE Virus IgG ELISA avidity

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

TBE Virus CSF/serum quotient IgG


sample type: serum
sample type 2: CSF
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

TBE virus CSF/serum quotient IgM

sample type: serum
sample type 2: CSF
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

TBE Virus IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA, IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21



Antibodies against

TBE Virus IgM

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA, IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Yellow fever virus IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: biweekly
reference value: see result report
remark: The analysis is not covered by the accreditation of our laboratory.
as at: 05/22

Yellow fever virus IgM


sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: biweekly
reference value: see result report
remark: The analysis is not covered by the accreditation of our laboratory.
as at: 05/22

Haemophilus influenzae IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
remark: The analysis is a LDT and not covered by the accreditation of our laboratory.
as at: 05/22

Haemophilus influenzae IgM

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
remark: The analysis is a LDT and not covered by the accreditation of our laboratory.
as at: 05/22



Antibodies against

Hanta-Profile-GLOBAL IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: EUROLINE
frequency: biweekly
reference value: see result report
as at: 12/21

Hanta-Profile-GLOBAL IgM

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: EUROLINE
frequency: biweekly
reference value: see result report
as at: 12/21

Hanta Virus Pool 1 IgG


sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: biweekly
reference value: see result report
as at: 12/21

Hanta Virus Pool 1 IgM

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: biweekly
reference value: see result report
as at: 12/21

Helicobacter pylori (CagA) IgA

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21



Antibodies against

Helicobacter pylori (CagA) IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Helicobacter pylori IgA

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA, EUROLINE/Westernblot
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Helicobacter pylori IgG


sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA, EUROLINE/Westernblot
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Hepatitis E IgA

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Hepatitis E IgAGM

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21



Antibodies against

Hepatitis E IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Hepatitis-E IgM

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

HHV-6 IgG


sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

HHV-6 IgM

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

HSV 1/2 Pool IgA

sample type: serum
sample type 2: EDTA plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21



Antibodies against

HSV 1/2 Pool IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

HSV 1/2 Pool IgM

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

HSV 1/2 Pool CSF/serum quotient IgG


sample type: serum
sample type 2: CSF
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

HSV C1, G1+G2 IgG EUROLINE/WB

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: EUROLINE/Westernblot
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

HSV C1, G1+G2 IgM EUROLINE/WB

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: EUROLINE/Westernblot
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21



Antibodies against

HSV-1 IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA, IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

HSV-1 IgM

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA, IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

HSV-2 IgG


sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA, IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

HSV-2 IgM

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA, IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Influenza A Virus IgA

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA, IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21



Antibodies against

Influenza A Virus IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA, IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Influenza A Virus IgM

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA, IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Influenza B Virus IgA


sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA, IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Influenza B Virus IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA, IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Influenza B Virus IgM

sample type: serum
sample type 2: EDTA plasma
sample volume: 1,5 ml
method: ELISA, IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21



Antibodies against

Japanese encephalitis virus IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: biweekly (Mo, Do)
reference value: see result report
as at: 12/21

Japanese encephalitis virus IgM

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: biweekly (Mo, Do)
reference value: see result report
as at: 12/21

Klebsiella pneumoniae IgG


sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
remark: The analysis is a LDT and not covered by the accreditation of our laboratory.
as at: 05/22

Klebsiella pneumoniae IgM

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
remark: The analysis is a LDT and not covered by the accreditation of our laboratory.
as at: 05/22

Leg. non-pneumophila IgAGM

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21



Antibodies against

Legionella pneumophila IgAGM

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Legionella pneumophila IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Legionella pneumophila ST1-7 IgA


sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Legionella pneumophila ST1-7 IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Legionella pneumophila ST1-7 IgM

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21



Antibodies against

Leishmania donovani IgA

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Leishmania donovani IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Leishmania donovani IgM


sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Listeria monocytogenes 1/2a IgA

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
remark: The analysis is a LDT and not covered by the accreditation of our laboratory.
as at: 05/22

Listeria monocytogenes 1/2a IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21



Antibodies against

Listeria monocytogenes 1/2a IgM

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Listeria monocytogenes 4b IgA

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
remark: The analysis is a LDT and not covered by the accreditation of our laboratory.
as at: 05/22

Listeria monocytogenes 4b IgG


sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Listeria monocytogenes 4b IgM

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Measles Virus IgG avidity

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21



Antibodies against

Measles Virus IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA, IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Measles Virus IgM

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA, IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Measles Virus CSF/serum quotient IgG


sample type: serum
sample type 2: CSF
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Mumps Virus CSF/serum quotient IgG

sample type: serum
sample type 2: CSF
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Mumps Virus IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA, ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21



Antibodies against

Mumps Virus IgM

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA, ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Mycoplasma hominis IgA

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Mycoplasma hominis IgG


sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Mycoplasma hominis IgM

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Mycoplasma pneumoniae IgA

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA, IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21



Antibodies against

Mycoplasma pneumoniae IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA, IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Mycoplasma pneumoniae IgM

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA, IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Parainfluenza Virus type 1-4 Pool IgA


sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Parainfluenza Virus type 1-4 Pool IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Parainfluenza Virus type 1-4 Pool IgM

sample type: serum
sample type 2: EDTA plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21



Antibodies against

Parvovirus B19 IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA, EUROLINE
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Parvovirus B19 IgM

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA, EUROLINE
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Plasmodium IgG


sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: biweekly
reference value: see result report
remark: This analysis is not intended to assess suitability for transfusion, transplantation or cell therapy.
as at: 05/22

RSV IgA

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA, IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

RSV IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA, IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21



Antibodies against

RSV IgM

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA, IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Rubella Virus IgG avidity

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA, IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Rubella Virus IFA avidity


sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA, IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Rubella Virus IgM

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Rubella Virus CSF/serum quotient IgG

sample type: serum
sample type 2: CSF
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21



Antibodies against

Sandfly Fever Virus IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
remark: The analysis is not covered by the accreditation of our laboratory.
as at: 05/22

Sandfly Fever Virus IgM

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
remark: The analysis is not covered by the accreditation of our laboratory.
as at: 05/22

SARS-CoV-2-NCP IgG


sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
sample type 3: dried blood spots (not covered by the accreditation of our laboratory)
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

SARS-CoV-2-NCP IgM

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
remark: The analysis is not covered by the accreditation of our laboratory.
as at: 05/22

SARS-CoV-2-Neutralisation assay

sample type: serum
sample type 2: EDTA, heparin plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21



Antibodies against

SARS-CoV-2-QuantiVac IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
sample type 3: dried blood spots (not covered by the accreditation of our laboratory)
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

SARS-CoV-2-S1-Spike-Protein IgA

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
remark:
as at: 12/21

SARS-CoV-2-S1-Spike-Protein IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
sample type 3: dried blood spots (not covered by the accreditation of our laboratory)
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Schistosoma mansoni IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA, ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Schistosoma mansoni IgM

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
remark: The analysis is not covered by the accreditation of our laboratory.
as at: 05/22



Antibodies against

Tetanus-Toxoid IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

TO.R.C.H. Profile IgG EUROLINE (CMV, HSV-1, HSV-2, Rubella, Toxoplasma gondii)

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: EUROLINE
frequency: daily (mo-fri)
reference value: see result report
remark: This analysis is not intended to assess suitability for transfusion, transplantation or cell therapy.
as at: 05/12

TO.R.C.H. Profile IgM EUROLINE (CMV, HSV-1, HSV-2, Rubella, Toxoplasma gondii)


sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: EUROLINE
frequency: daily (mo-fri)
reference value: see result report
remark: This analysis is not intended to assess suitability for transfusion, transplantation or cell therapy.
as at: 05/22

Toxoplasma gondii IgA

sample type: serum
sample type 2: EDTA plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Toxoplasma gondii IgG avidity

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA, IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21



Antibodies against

Toxoplasma gondii IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA, IFA
frequency: daily (mo-fri)
reference value: see result report
remark: This analysis is not intended to assess suitability for transfusion, transplantation or cell therapy.
as at: 05/22

Toxoplasma gondii IgM

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA, IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Toxoplasma gondii CSF/serum quotient IgG


sample type: serum
sample type 2: CSF
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Treponema pallidum IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA, IFA, EUROLINE
frequency: daily (mo-fri)
reference value: see result report
remark: This analysis is not intended to assess suitability for transfusion, transplantation or cell therapy.
as at: 05/22

Treponema pallidum IgGM

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
remark: This analysis is not intended to assess suitability for transfusion, transplantation or cell therapy.
as at: 05/22



Antibodies against

Treponema pallidum IgM

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA, IFA, EUROLINE
frequency: daily (mo-fri)
reference value: see result report
remark: This analysis is not intended to assess suitability for transfusion, transplantation or cell therapy.
as at: 05/22

Treponema pallidum CSF/serum quotient IgG

sample type: serum
sample type 2: CSF
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Ureaplasma urealyticum IgG


sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Ureaplasma urealyticum IgM

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Usutu Virus IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,0 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
remark: The analysis is not covered by the accreditation of our laboratory.
as at: 05/22



Antibodies against

VZV IgA

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA, IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

VZV IgG avidity

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA, IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

VZV IgG


sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA, IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

VZV IgM

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA, IFA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

VZV CSF/serum quotient IgG

sample type: serum
sample type 2: CSF
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21



Antibodies against

West Nile Virus IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: biweekly
reference value: see result report
as at: 12/21

West Nile Virus IgM

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: biweekly
reference value: see result report
as at: 12/21

Yersinia enterocolitica IgA


sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA, IFA, Westernblot
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Yersinia enterocolitica IgG

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA, IFA, Westernblot
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Zika IgA

sample type: serum
sample type 2: EDTA, heparin plasma
sample volume: 1,5 ml
method: ELISA
frequency: biweekly
reference value: see result report
as at: 12/21



Antibodies against

Zika IgAM

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: biweekly
reference value: see result report
as at: 12/21

Zika Virus IgG

sample type: serum
sample type 2: EDTA, heparin plasma
sample volume: 1,5 ml
method: ELISA
frequency: biweekly
reference value: see result report
as at: 12/21

Zika Virus IgG


sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: IFA
frequency: biweekly
reference value: see result report
remark: The analysis is not covered by the accreditation of our laboratory.
as at: 05/22

Zika Virus IgM

sample type: serum
sample type 2: EDTA, heparin plasma
sample volume: 1,5 ml
method: ELISA
frequency: biweekly
reference value: see result report
as at: 12/21

Zika Virus IgM

sample type: serum
sample type 2: EDTA-, Heparin-, Citrat-Plasma
sample volume: 1,5 ml
method: IFA
frequency: biweekly
reference value: see result report
remark: The analysis is not covered by the accreditation of our laboratory.
as at: 05/22



Antibodies class IgE against (allergy diagnostics)

total IgE

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Profile Atopy 13

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: EUROLINE
frequency: daily (mo-thu)
reference value: see result report
remark: allergens see test order form
remark 2: The analysis is not covered by the accreditation of our laboratory.
as at: 05/22

Profile Atopy Screen


sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: EUROLINE
frequency: daily (mo-thu)
reference value: see result report
remark: allergens see test order form
remark 2: The analysis is not covered by the accreditation of our laboratory.
as at: 05/22

Profile Atopy

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: EUROLINE
frequency: daily (mo-thu)
reference value: see result report
remark: allergens see test order form
as at: 12/21

Profile DPA-Dx insect venoms 3 IgE

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 2 ml
method: EUROASSAY
frequency: daily (mo-fri)
reference value: see result report
remark: allergens see test order form
as at: 12/21



Antibodies class IgE against (allergy diagnostics)

Profile DPA-Dx insect venoms 3

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: EUROLINE
frequency: daily (mo-fri)
reference value: see result report
as at: 02/22

Profile DPA-Dx Pollen 1 IgE

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 2 ml
method: EUROASSAY
frequency: daily (mo-fri)
reference value: see result report
remark: allergens see test order form
as at: 12/21

Profile DPA-Dx pollen 1


sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: EUROLINE
frequency: daily (mo-fri)
reference value: see result report
as at: 12/21

Profile DPA-Dx Paediatrics 1 IgE

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 2 ml
method: EUROASSAY
frequency: daily (mo-fri)
reference value: see result report
remark: allergens see test order form
as at: 12/21

Profile Atopy Screen_2

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: EUROLINE
frequency: daily (mo-thu)
reference value: see result report
remark: allergens see test order form
remark 2: The analysis is not covered by the accreditation of our laboratory.
as at: 05/22



Antibodies class IgE against (allergy diagnostics)

Profile Inhalation 2 IgE

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 2 ml
method: EUROASSAY
frequency: daily (mo-fri)
reference value: see result report
remark: allergens see test order form
as at: 12/21

Profile Inhalation 2

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: EUROLINE
frequency: daily (mo-thu)
reference value: see result report
remark: allergens see test order form
as at: 12/21

Profil Inhalation_2


sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: EUROLINE
frequency: daily (mo-thu)
reference value: see result report
remark: allergens see test order form
remark 2: The analysis is not covered by the accreditation of our laboratory.
as at: 05/22

Profile Inhalation IgE

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 2 ml
method: EUROASSAY
frequency: daily (mo-fri)
reference value: see result report
remark: allergens see test order form
as at: 12/21

Profil Inhalation

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: EUROLINE
frequency: daily (mo-thu)
reference value: see result report
remark: allergens see test order form
as at: 12/21



Antibodies class IgE against (allergy diagnostics)

Profile Food 2 IgE

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 2 ml
method: EUROASSAY
frequency: daily (mo-fri)
reference value: see result report
remark: allergens see test order form
as at: 12/21

Profile Food 2

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: EUROLINE
frequency: daily (mo-thu)
reference value: see result report
remark: allergens see test order form
as at: 12/21

Food profile 3


sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: EUROLINE
frequency: daily (mo-thu)
reference value: see result report
remark: allergens see test order form
as at: 12/21

Profile Food IgE

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 2 ml
method: EUROASSAY
frequency: daily (mo-fri)
reference value: see result report
remark: allergens see test order form
as at: 12/21

Profile Food

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: EUROLINE
frequency: daily (mo-thu)
reference value: see result report
remark: allergens see test order form
remark 2: The analysis is not covered by the accreditation of our laboratory.
as at: 05/22



Antibodies class IgE against (allergy diagnostics)

Profile Paediatric / Atopy 2 IgE

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 2 ml
method: EUROASSAY
frequency: daily (mo-fri)
reference value: see result report
remark: allergens see test order form
as at: 12/21

Profile Paediatric / Atopy IgE


sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 2 ml
method: EUROASSAY
frequency: daily (mo-fri)
reference value: see result report
remark: allergens see test order form
as at: 12/21

Profile paediatrics

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: EUROLINE
frequency: daily (mo-thu)
reference value: see result report
remark: allergens see test order form
as at: 12/21

Profile Paediatric inhalation

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: EUROLINE
frequency: daily (mo-thu)
reference value: see result report
remark: allergens see test order form
as at: 12/21



25-OH-vitamin D ELISA

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5ml
method: ELISA
frequency: biweekly
reference value: see result report
remark: The analysis is not covered by the accreditation of our laboratory.
as at: 12/21

APOE

sample type: EDTA blood
sample type2: isolated genomic DNA
sample volume: 1,5 ml
method: EUROARRAY
frequency: biweekly
reference value: see result report
remark: Declaration of consent for human genetic diagnostics required.
remark 2: The analysis is not covered by the accreditation of our laboratory.
as at: 12/21

Calcitonin hormone assay


sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: biweekly
reference value: see result report
remark: The analysis is not covered by the accreditation of our laboratory.
as at: 12/21

Cortisol Saliva

sample type: saliva
sample volume: 1,5ml
method: ELISA
frequency: biweekly
reference value: see result report
remark: The analysis is not covered by the accreditation of our laboratory.
as at: 01/19

dermatomycosis

sample type: skin material
sample type 2: hair material
sample type 3: nail material
sample type 4: culture material
sample type 5: purified DNA
sample volume: culture material 1 cm²
method: EUROARRAY
frequency: biweekly
reference value: see result report
remark: The analysis is not covered by the accreditation of our laboratory.
as at: 12/21



FII+

sample type: EDTA blood
sample type 2: isolated genomic DNA
sample volume: 1,5 ml
method: EUROARRAY
frequency: biweekly
reference value: see result report
remark: Declaration of consent for human genetic diagnostics required.
remark 2: The analysis is not covered by the accreditation of our laboratory.
as at: 12/21

FV

sample type: EDTA blood
sample type2: isolated genomic DNA
sample volume: 1,5 ml
method: EUROARRAY
frequency: biweekly
reference value: see result report
remark: Declaration of consent for human genetic diagnostics required.
remark 2: The analysis is not covered by the accreditation of our laboratory.
as at: 12/21

HLA-B27


sample type: EDTA blood
sample volume: 1,5 ml
method: EUROARRAY
frequency: biweekly
reference value: see result report
remark: Declaration of consent for human genetic diagnostics required.
as at: 12/21

HLA-B57

sample type: EDTA blood
sample volume: 1,5 ml
method: EUROARRAY
frequency: biweekly
reference value: see result report
remark: Declaration of consent for human genetic diagnostics required.
as at: 12/21

HLA-Cw6

sample type: EDTA blood
sample volume: 1,5 ml
method: EUROARRAY
frequency: biweekly
reference value: see result report
remark: Declaration of consent for human genetic diagnostics required.
as at: 12/21



HLA-DQ2/DQ8

sample type: EDTA blood
sample volume: 1,5 ml
method: EUROARRAY
frequency: biweekly
reference value: see result report
remark: Declaration of consent for human genetic diagnostics required.
as at: 12/21

HLA-DRB1 shared epitopes

sample type: EDTA blood
sample type2: isolated genomic DNA
sample volume: 1,5 ml
method: EUROARRAY
frequency: biweekly
reference value: see result report
remark: Declaration of consent for human genetic diagnostics required.
remark 2: The analysis is not covered by the accreditation of our laboratory.
as at: 12/21

HPV


sample type: cervical swab
sample volume: n/a
method: EUROARRAY
frequency: biweekly
reference value: see result report
remark: The analysis is not covered by the accreditation of our laboratory.
as at: 12/21

Haemochromatosis (HFE) (2 SNP+)

sample type: EDTA blood
sample type2: isolated genomic DNA
sample volume: 1,5 ml
method: EUROARRAY
frequency: biweekly
reference value: see result report
remark: Declaration of consent for human genetic diagnostics required.
remark 2: The analysis is not covered by the accreditation of our laboratory.
as at: 12/21

Haemochromatosis (HFE) (4 SNP+)

sample type: EDTA blood
sample type2: isolated genomic DNA
sample volume: 1,5 ml
method: EUROARRAY
frequency: biweekly
reference value: see result report
remark: Declaration of consent for human genetic diagnostics required.
remark 2: The analysis is not covered by the accreditation of our laboratory.
as at: 12/21



Intact PTH hormone assay

sample type: serum
sample type 2: EDTA, heparin, citrat plasma
sample volume: 1,5 ml
method: ELISA
frequency: biweekly
reference value: see result report
remark: The analysis is not covered by the accreditation of our laboratory.
as at: 12/21

Lactose intolerance

sample type: EDTA blood
sample type2: isolated genomic DNA
sample volume: 1,5 ml
method: EUROARRAY
frequency: biweekly
reference value: see result report
remark: Declaration of consent for human genetic diagnostics required.
remark 2: The analysis is not covered by the accreditation of our laboratory.
as at: 12/21

MTHFR


sample type: EDTA blood
sample type2: isolated genomic DNA
sample volume: 1,5 ml
method: EUROARRAY
frequency: biweekly
reference value: see result report
remark: Declaration of consent for human genetic diagnostics required.
remark 2: The analysis is not covered by the accreditation of our laboratory.
as at: 12/21

Mullerian-duct repression hormone (MRH)

sample type: serum
sample type 2: heparin plasma
sample volume: 1,5 ml
method: ELISA
frequency: daily (mo-fri)
reference value: see result report
remark: The analysis is not covered by the accreditation of our laboratory.
as at: 01/19

SARS-CoV-2

sample type: throat swab
sample type 2: purified RNA
sample volume: -
method: EURORealTime
frequency: biweekly
reference value: see result report
as at: 12/21



sCD163

sample type: urine
sample volume: 1,5ml
method: ELISA
frequency: biweekly
reference value: see result report
remark: The analysis is not covered by the accreditation of our laboratory.
as at: 10/19

STI-11

sample type: swab
sample type 2: urine
sample volume: 1,5 ml
method: EUROARRAY
frequency: biweekly
reference value: see result report
remark: The analysis is not covered by the accreditation of our laboratory.
as at: 12/21

STI-6


sample type: swab
sample type 2: urine
sample volume: 1,5 ml
method: EUROARRAY
frequency: biweekly
reference value: see result report
remark: The analysis is not covered by the accreditation of our laboratory.
as at: 12/21

STI-7

sample type: swab
sample type 2: urine
sample volume: 1,5 ml
method: EUROARRAY
frequency: biweekly
reference value: see result report
remark: The analysis is not covered by the accreditation of our laboratory.
as at: 12/21

STI-CT/NG

sample type: swab
sample type 2: urine
sample volume: 1,5 ml
method: EUROARRAY
frequency: biweekly
reference value: see result report
remark: The analysis is not covered by the accreditation of our laboratory.
as at: 12/21



STI-CT/NG/TP/TV

sample type: swab
sample type 2: urine
sample volume: 1,5 ml
method: EUROARRAY
frequency: biweekly
reference value: see result report
remark: The analysis is not covered by the accreditation of our laboratory.
as at: 12/21

STI-HSV1/2

sample type: swab
sample type 2: urine
sample type 3: mouth/pharyngeal lavage
sample volume: 1,5 ml
method: EUROARRAY
frequency: biweekly
reference value: see result report
remark: The analysis is not covered by the accreditation of our laboratory.
as at: 08/2020

STI

sample type: swab
sample type 2: urine
sample volume: 1,5 ml
method: EUROARRAY
frequency: biweekly
reference value: see result report
remark: The analysis is not covered by the accreditation of our laboratory.
as at: 12/21

Uromodulin

sample type: serum
sample type 2: heparin plasma
sample volume: 1,5 ml
method: ELISA
frequency: biweekly
reference value: see result report
remark: The analysis is not covered by the accreditation of our laboratory.
as at: 01/19